



Genetex GTX27842 RRID:AB_371884 GP INS on Mouse Tissue (Wright Lab)


Positive control Slide
(Cell Marque GP Insulin 1:500)



Genetex GTX27842 GP INS
1:500



Genetex GTX27842 GP INS
1:1000



Insulin
Dapi

Identical slides were run in the same experiment; only difference between the slides was the primary antibody. Two different lot number tubes of GTX27842 contained a contaminating particulates.